Questions and Answers on RFP NNG13477846R

Question 1:   Reviewing the documentation for solicitation NNG13477846R and had a question regarding Appendix A. Item number two in the critical requirements table specifically the eLISA/NGO column seems to be missing a value. Are the three blocks in front of RMS supposed to be a numerical value or are you looking for the vendor to state what is achievable?

The three blocks in front of RMS should read λ/30. The full specification is that the wavefront quality must be better than or equal to λ/30 RMS.